

## Vapor Immobilization Sling

Warnings and Instructions: Review carefully, proper application is required Marning: This device will not prevent or eliminate risk of injury. Do Not Overtighten. If swelling, pain, skin irritation, or an unusual reaction occurs, discontinue use immediately and consult your medical professional. Care: Hand wash using mild soap. Rinse thoroughly. Air dry only. Do not tumble dry.



1. Place arm in the sling pouch.



3. Pull waist strap around waist and clip into clip at front of brace.



5. Make final adjustments with straps at front of brace.



2. Attach shoulder strap hook and loop to corners of brace. Secure top of brace with elbow strap.




4. Adjust straps at back of brace to provide proper fit.



6. Place thumb in thumb strap for comfort.

## Vapor Immobilization Sling


## Warnings and Instructions: Review carefully, proper application is required Warning: This device will not prevent or eliminate risk of injury. Do Not Overtighten. If swelling, pain, skin irritation, or an unusual reaction occurs, discontinue use immediately and consult your medical professional. Care: Hand wash using mild soap. Rinse thoroughly. Air dry only. Do not tumble dry.



1. Place arm in the sling pouch.



3. Pull waist strap around waist and clip into clip at front of brace.



5. Make final adjustments with straps at front of brace.



2. Attach shoulder strap hook and loop to corners of brace. Secure top of brace with elbow strap.



4. Adjust straps at back of brace to provide proper fit.



6. Place thumb in thumb strap for comfort.



1185 E Main St., Santa Paula, California 93060 California: 800-221-5465, National: 800-654-3241 International: 805-525-4244, fax: 805-933-2348 U.S. fax: 800-559-5975, www.hely-weber.com



Rev B 2/14



1185 E Main St., Santa Paula, California 93060 California: 800-221-5465, National: 800-654-3241 International: 805-525-4244, fax: 805-933-2348 U.S. fax: 800-559-5975, www.hely-weber.com





504