
Twin V Low Profile with Sagittal Panels

Warnings and Instructions: Review carefully, proper application is required

Warning: This device will not prevent or eliminate risk of injury. Do Not Overtighten, If swelling, pain, skin irritation, or an unusual reaction occurs, discontinue use immediately and consult your medical professional. Care: Hand wash using mild soap. Rinse thoroughly. Air dry only. Do not tumble dry.



1. Release side pull straps, and position the brace with the back posterior panel centered over the lumbar area.



3. Center brace. The front plastic panel should be centered over the abdomen.



5. Finished application.



View Application Video



1185 E Main St., Santa Paula, California 93060 California: 800-221-5465, National: 800-654-3241 International: 805-525-4244, fax: 805-933-2348 U.S. fax: 800-559-5975, www.hely-weber.com




419


Twin V Low Profile with Sagittal Panels

Warnings and Instructions: Review carefully, proper application is required

Warning: This device will not prevent or eliminate risk of injury. Do Not Overtighten, If swelling, pain, skin irritation, or an unusual reaction occurs, discontinue use immediately and consult your medical professional. Care: Hand wash using mild soap. Rinse thoroughly. Air dry only. Do not tumble dry.




1. Release side pull straps, and position the brace with the back posterior panel centered over the lumbar area.




2. Stretch out right side of brace to its furthest extension. Snugly secure right panel over left panel.





3. Center brace. The front plastic panel should be centered over the abdomen.





6.






1185 E Main St., Santa Paula, California 93060 California: 800-221-5465, National: 800-654-3241 International: 805-525-4244, fax: 805-933-2348 U.S. fax: 800-559-5975, www.hely-weber.com

E EC REP MDSS GmbH, Schiffgraben 41 30175 Hannover, Germany

> 4/15 Rev C



5. Finished application.











6

© 2017 Weber Orthopedics Inc. DBA Helv & Weber all rights reserved

6. Finished application back.

2. Stretch out right side of brace to its furthest exten-

sion. Snugly secure right panel over left panel.

4. Tighten the pull straps and secure.