


# Clavicle Splint


661

## Warnings and Instructions: Review carefully, proper application is required


**⚠ Warning:** This device will not prevent or eliminate risk of injury. Do Not Overtighten. If swelling, pain, skin irritation, or an unusual reaction occurs, discontinue use immediately and consult your medical professional.  
Care: Hand wash using mild soap. Rinse thoroughly. Air dry only. Do not tumble dry.




1. Position the D-ring pad in the center of the back. Place straps over each shoulder.




2. Secure the brace on the unaffected side first. Slide the strap through the buckle. Repeat for the affected side.



3. Adjust the affected side until the shoulders are in normal anatomical position.



4. Provide the final tension by simultaneous strap adjustment. Patient should be in a resting position with the shoulders back. **DO NOT HYPEREXTEND THE PATIENT.**




# Clavicle Splint


661

## Warnings and Instructions: Review carefully, proper application is required


**⚠ Warning:** This device will not prevent or eliminate risk of injury. Do Not Overtighten. If swelling, pain, skin irritation, or an unusual reaction occurs, discontinue use immediately and consult your medical professional.  
Care: Hand wash using mild soap. Rinse thoroughly. Air dry only. Do not tumble dry.




1. Position the D-ring pad in the center of the back. Place straps over each shoulder.



2. Secure the brace on the unaffected side first. Slide the strap through the buckle. Repeat for the affected side.



3. Adjust the affected side until the shoulders are in normal anatomical position.



4. Provide the final tension by simultaneous strap adjustment. Patient should be in a resting position with the shoulders back. **DO NOT HYPEREXTEND THE PATIENT.**

